

Volumetric flasks with 5000 ml cap

Flasks with calibration ring

with ground cone NS

and cap in PE polyethylene with octagonal handle.


Tolerances within the regulations for Class A

Borosilicate glass 3.3

Calibrated "In" (to contain)

DIN-EN-ISO 1042

On request are available volumetric flasks in Class A with Batch Certificate of Conformity (ref. Codes 12582xx)



Code	Description	Capacity ml	Accuracy ml	NS cone	Pieces x conf.
1258173	Volumetric flasks with 5000 ml cap	5000	1.200	34/35	1