

Non-sterile 5 ml serological graduated pipettes

div. 1/10

Pipettes made of crystal polystyrene, transparent and biologically inert, in a single piece without welding, with safety pad in cellulose acetate.



- Inaccuracy: $<\pm 1\%$
- Imprecision: $<2\%$
- Sterilization: ionizing radiation in accordance with a method validated in accordance with ISO 11137-1: 2006.
- Apyrogenicity: absence of endotoxins, therefore free from pyrogenic substances due to the particular construction process.
- DNase - RNase: due to the highly automated production and packaging method with which they are produced, pipettes normally lack these enzymes.
- Color code: compliant with ISO 1769.
- In Vitro Diagnostic Medical Devices (IVD) in compliance with which they bear the CE marking

Code	Description	Capacity ml	Color	Single pack	Sterile	Pieces per pack.
02160519	Non-sterile 5 ml serological graduated pipettes	5	blue	no	no	500