

## Graduated serological pipettes ml 25 sterile in single pack

### div. 2/10

Pipettes made of crystal polystyrene, transparent and biologically inert, in a single piece without welding, with safety pad in cellulose acetate.



- Inaccuracy:  $\leq \pm 1\%$
- Imprecision:  $\leq 2\%$
- Sterilization: ionizing radiation in accordance with a method validated in accordance with ISO 11137-1: 2006.
- Apyrogenicity: absence of endotoxins, therefore free from pyrogenic substances due to the particular construction process.
- DNase - RNase: due to the highly automated production and packaging method with which they are produced, pipettes normally lack these enzymes.
- Color code: compliant with ISO 1769.
- In Vitro Diagnostic Medical Devices (IVD) in compliance with which they bear the CE marking

Code	Description	Capacity ml	Color	Single pack	Sterile	Pieces per pack.
02162510	Graduated serological pipettes ml 25 sterile in single pack	25	red	yes	yes	200